

HBsAb Hepatitis B Surface Antibody Rapid Diagnostic Kit (Immunochromatography)

Product Name

HBsAb Hepatitis B Surface Antibody Rapid Test Kit(Immunochromatography)

Intended Use

Used for the qualitative detection of Hepatitis B surface antibody in serum/plasma.

Test Principle

The sample mixing up colloidal-gold monoclonal antibody move along the membrane to the T line, and form the T line when the human serum/plasma contains HBsAb according to the principle of double antibody sandwich method and gold immunochromatography assay, which is a positive result. Unreacted markers move forward continually to combine with anti-mouse antibody and form a control line. If the test line does not appear, it is a negative result.

Main Components

Sample pad, colloidal gold marked pad, nitrocellulose membrane, absorbent paper and PVC board.

Storage and Expiry

Store as packaged in the sealed pouch at 4-30°C, avoid hot and sunshine, dry place, valid for 24 months. DO NOT FREEZE. Some protective measures should be taken in hot summer and cold winter to avoid high temperature or freeze-thaw.

Sample Requirement

1. Collect venous blood a clean and dry container according to the standard method. Separate the serum or plasma for testing. EDTA, sodium citrate, sodium oxalate, heparin can be used as the anticoagulants
2. After collection, separate the serum and plasma as soon as possible, in order to avoid hemolysis.
3. Use fresh samples. Samples may be stored at 2-8°C for 3 days, and should be stored at -20 °C if cannot be tested immediately. Do not freeze and thaw the sample repeatedly.

Test Procedure

Instructions must be read entirely before taking the test. Allow the test device controls to equilibrate to room temperature for 30 minutes (20°C-30°C) prior to testing. Do not open the inner packaging until ready, it must be used in one hour if opened (Humidity: 20%~90%, Temp: 10°C-50°C)

Strip:1. Remove the test device from the sealed pouch, put the end of the test strip print with arrow into the sample (about 50-60μl), the interface of serum/plasma should not exceed the max line, take it out and place the test device on a clean and level surface after 10 seconds. 2.Observe

the test result immediately within 15~20 minutes, the result is invalid over 20 minutes.

Cassette: 1. Take off the outer packing, put the cassette onto the desk with the sample window up.

2. Drop 3 drops of serum or plasma (80μl-100μl) vertically into the sample hole of cassette.


3. Observe the test results immediately within 15~20 minutes, the result is invalid over 20 minutes.

Result Judgment

POSITIVE: Two distinct red lines appear. One line should be in the control region (C) and the other line should be in the test region (T).

NEGATIVE: One red line appears in the control region(C). No apparent red or pink line appears in the test region (T).

INVALID: No red lines appear or control line fails to appear, indicating that the operator error or reagent failure. Verify the test procedure and repeat the test with a new testing device.



Limitation

1. This reagent is designed for the qualitative screening test. Concentration of HBsAb cannot be determined by this qualitative test.
2. The results of the reagent are only for clinical reference, a confirmed diagnosis should be made only by a physician after all clinical and laboratory findings have been evaluated.

Precaution

1. The test result is invalid over 20 minutes.
2. Do not use after the expiration.
3. The strength of the quality control line doesn't indicate the quality problem of the reagent, a test result that is clearly visible demonstrates the reagent is effective.
4. Both deep and light color may occur during the test, even a very light brand during observation time should be judged as positive a result.
5. Do not use other kinds of quality control sample to test the reagent. Components of different batches cannot be exchanged for use to avoid erroneous results.